

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 08/30/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:46
INTERNAL STANDARD	VERIFIED	22:46
REFERENCE STANDARD TEMP	33.94c	
REFERENCE STANDARD	0.09	22:46
BLANK TEST	0.00	22:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:47
BLANK TEST	0.00	22:48
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	22:48
BLANK TEST	0.00	22:49



SGS High / Butters WCSO SHIACD #162 08-30-10