

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/06/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:54
INTERNAL STANDARD	VERIFIED	02:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:56
BLANK TEST	0.00	02:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:57



SGT *Ralph Bartlett* WCSO #142 09-06-10