

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/14/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:39
INTERNAL STANDARD	VERIFIED	03:39
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:40
BLANK TEST	0.00	03:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:41
BLANK TEST	0.00	03:42
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	03:42
BLANK TEST	0.00	03:43



SOS *Ralph Bartlett*

WCSO #162

09-14-10