

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 09/20/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10110

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:50 |
| INTERNAL STANDARD | VERIFIED | 00:50 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 00:50 |
| BLANK TEST | 0.00 | 00:51 |
| REFERENCE STANDARD TEMP | 33.97c | |
| REFERENCE STANDARD | 0.09 | 00:52 |
| BLANK TEST | 0.00 | 00:52 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 00:53 |
| BLANK TEST | 0.00 | 00:53 |



505 Ralph J Bartlett

09-20-10

WCSO #166