

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 09/29/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:19
INTERNAL STANDARD	VERIFIED	01:19
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	01:20
BLANK TEST	0.00	01:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:21
BLANK TEST	0.00	01:22
REFERENCE STANDARD TEMP	33.93c	
REFERENCE STANDARD	0.09	01:22
BLANK TEST	0.00	01:23



505 Rully B... wess #162 9/29/10