

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/07/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:32
INTERNAL STANDARD	VERIFIED	01:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:33
BLANK TEST	0.00	01:33
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	01:34
BLANK TEST	0.00	01:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:35
BLANK TEST	0.00	01:35



SCS Run 1 Run 10/7/10 SHERIFF #162