

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/21/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:17
INTERNAL STANDARD	VERIFIED	04:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:18
BLANK TEST	0.00	04:18
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:19
BLANK TEST	0.00	04:19
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:20
BLANK TEST	0.00	04:20



SOT Ralph Bartlett

10/21/10

WCSO # 102