

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/01/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:53
INTERNAL STANDARD	VERIFIED	02:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:53
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:54
BLANK TEST	0.00	02:55
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:55
BLANK TEST	0.00	02:56



505 Rugh & Burtch weso SHERID #162 11-01-10