

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/09/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:45
INTERNAL STANDARD	VERIFIED	00:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:45
BLANK TEST	0.00	00:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	00:46
BLANK TEST	0.00	00:47
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	00:47
BLANK TEST	0.00	00:48



505 *Ralph J Bartlett*

11-09-10

WCSD # 162