

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/16/2010

Operator Name: CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:38
INTERNAL STANDARD	VERIFIED	15:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	15:38
BLANK TEST	0.00	15:39
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	15:40
BLANK TEST	0.00	15:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:41
BLANK TEST	0.00	15:41

*Charles B. Geisler #248*

