

**SUPERVISOR TEST**

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/24/2010

Operator Name: CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:32
INTERNAL STANDARD	VERIFIED	11:33
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	11:33
BLANK TEST	0.00	11:34
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	11:34
BLANK TEST	0.00	11:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:35
BLANK TEST	0.00	11:36

*Charles B. Geisler #248*

