

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/03/2010

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:00
INTERNAL STANDARD	VERIFIED	13:00
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	13:01
BLANK TEST	0.00	13:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:02
BLANK TEST	0.00	13:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:03
BLANK TEST	0.00	13:04

So. Ill. D. #248

