

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/26/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10190

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 02:07 |
| INTERNAL STANDARD | VERIFIED | 02:07 |
| REFERENCE STANDARD TEMP | 33.96c | |
| REFERENCE STANDARD | 0.09 | 02:07 |
| BLANK TEST | 0.00 | 02:08 |
| REFERENCE STANDARD TEMP | 33.94c | |
| REFERENCE STANDARD | 0.09 | 02:08 |
| BLANK TEST | 0.00 | 02:09 |
| REFERENCE STANDARD TEMP | 33.93c | |
| REFERENCE STANDARD | 0.09 | 02:09 |
| BLANK TEST | 0.00 | 02:10 |



SGT Ralph Bartlett WCSO #162 12-26-10