

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 01/01/2010

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:19
INTERNAL STANDARD	VERIFIED	01:19
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:20
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:21
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:22
BLANK TEST	0.00	01:23



567 *Ralph Bartlett* WOSO #162 01-01-10