

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/08/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:36
INTERNAL STANDARD	VERIFIED	08:36
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	08:37
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	08:38
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	08:39
BLANK TEST	0.00	08:40



SGT Russ Burton 1/08/10