

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/14/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:37
INTERNAL STANDARD	VERIFIED	03:37
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:37
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:38
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:39
BLANK TEST	0.00	03:40



SGT *Ralph J Bartlett* SITE # 162 1-14-10