

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/28/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09260

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:30
INTERNAL STANDARD	VERIFIED	03:30
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:30
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:31
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:32
BLANK TEST	0.00	03:33



505 *Ralph Bartlett* W056 #162 1-28-10