

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/17/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:17
INTERNAL STANDARD	VERIFIED	02:17
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:17
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:19
BLANK TEST	0.00	02:20



SGS Ralph J Bartlett WCO SHERIFF #162 2-17-10