

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/25/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:40
INTERNAL STANDARD	VERIFIED	02:40
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:41
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	02:42
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	02:43
BLANK TEST	0.00	02:43



See page 1 Report 2/25/10 WESO SHERIFF #162