

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/02/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:32
INTERNAL STANDARD	VERIFIED	00:32
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:32
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:33
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:34
BLANK TEST	0.00	00:35



SGT Ralph Bartlett #102 3-2-10