

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/08/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:14
INTERNAL STANDARD	VERIFIED	04:14
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:14
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	04:15
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	04:16
BLANK TEST	0.00	04:17



SGS *Ralph Bartlett*

WC SO SHERIFF #162

03-08-10