

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/16/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:25
INTERNAL STANDARD	VERIFIED	03:25
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:25
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	03:26
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	03:27
BLANK TEST	0.00	03:28



SGT *Ralph Bartlett*

WARREN CO SHERIFF #162

03-16-10