

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 03/25/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 09310

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:17
INTERNAL STANDARD	VERIFIED	01:17
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:17
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:18
REFERENCE STANDARD TEMP	33.9c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:20

