

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/31/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:06
INTERNAL STANDARD	VERIFIED	01:06
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:06
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:07
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:08
BLANK TEST	0.00	01:09

