

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/08/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:18
INTERNAL STANDARD	VERIFIED	00:18
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	00:19
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	00:20
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	00:21
BLANK TEST	0.00	00:21



SG- Russ, Bartlett wase #162 04-08-10