

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/22/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:59
INTERNAL STANDARD	VERIFIED	22:59
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	23:00
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	23:01
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	23:02
BLANK TEST	0.00	23:03



SCS Russy Burtos waco #162 4/22/10