

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/02/2010

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:24
INTERNAL STANDARD	VERIFIED	09:25
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	09:25
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	09:26
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	09:27
BLANK TEST	0.00	09:28

for Charles B. Geisler #210

