

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/08/2010

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:42
INTERNAL STANDARD	VERIFIED	14:43
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	14:43
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	14:44
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	14:45
BLANK TEST	0.00	14:46

for C.B. Geisler #448

