

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/19/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10010

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:24
INTERNAL STANDARD	VERIFIED	17:24
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	17:25
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	17:26
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.10	17:27
BLANK TEST	0.00	17:27



505 *Ralph Bartlett* 5/19/10