

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/27/2010

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10010

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 10:44 |
| INTERNAL STANDARD | VERIFIED | 10:44 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 10:44 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 10:45 |
| REFERENCE STANDARD TEMP | 34.0c | |
| REFERENCE STANDARD | 0.09 | 10:46 |
| BLANK TEST | 0.00 | 10:47 |

Charles B. Geisler
#218

