

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/01/2010

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:14
INTERNAL STANDARD	VERIFIED	11:14
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	11:15
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	11:16
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	11:17
BLANK TEST	0.00	11:18



Handwritten signature
#248 6/1/10 1120