

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/26/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:25
INTERNAL STANDARD	VERIFIED	23:25
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	23:25
REFERENCE STANDARD TEMP	33.9c	
REFERENCE STANDARD	0.10	23:26
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	23:27
BLANK TEST	0.00	23:28



SGT *Ralph Bartlett*

06/27/10