

SOLUTION CHANGE




 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 07/01/2010

Operator Name: RALPH J BARTLETT
 Permit #: 26222
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:25
INTERNAL STANDARD	VERIFIED	01:25
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:25
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:26
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	01:27
BLANK TEST	0.00	01:28



SG5 Rusky Purvey 07/01/10