

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/07/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:39
INTERNAL STANDARD	VERIFIED	00:39
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:40
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:41
REFERENCE STANDARD TEMP	34.0c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:42



SOI *Ralph Bartlett* 7/7/10