

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 07/15/2010

Operator Name: JOSHUA M LOPEZ  
Permit #: 34728  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:47
INTERNAL STANDARD	VERIFIED	15:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	15:47
BLANK TEST	0.00	15:48
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	15:48
BLANK TEST	0.00	15:49
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	15:49
BLANK TEST	0.00	15:50



SGS Rep of Results 07-15-10