

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dnt: 144506

Date: 07/21/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10060

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:12
INTERNAL STANDARD	VERIFIED	23:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:13
BLANK TEST	0.00	23:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	23:14
BLANK TEST	0.00	23:14
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	23:15
BLANK TEST	0.00	23:15



SGT Rush / Bartlett WJSG #162 07-25-10