

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/03/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:39
INTERNAL STANDARD	VERIFIED	02:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	33.96c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42



505 *Ralph Bartlett*

WCSO #162

08-03-10