

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/09/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10110

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:54 |
| INTERNAL STANDARD | VERIFIED | 01:54 |
| REFERENCE STANDARD TEMP | 33.99c | |
| REFERENCE STANDARD | 0.09 | 01:54 |
| BLANK TEST | 0.00 | 01:55 |
| REFERENCE STANDARD TEMP | 34.04c | |
| REFERENCE STANDARD | 0.09 | 01:55 |
| BLANK TEST | 0.00 | 01:56 |
| REFERENCE STANDARD TEMP | 33.98c | |
| REFERENCE STANDARD | 0.09 | 01:56 |
| BLANK TEST | 0.00 | 01:57 |



SGT *Ralph Bartlett*

WCSO #162

8/9/10