

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/17/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:33
INTERNAL STANDARD	VERIFIED	02:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:33
BLANK TEST	0.00	02:34
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	02:35
BLANK TEST	0.00	02:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:36



SGT *Ralph Bartlett* 8/17/10 WCSO #142