

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 08/30/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10119

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:00
INTERNAL STANDARD	VERIFIED	00:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:01
BLANK TEST	0.00	00:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:02
BLANK TEST	0.00	00:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:03
BLANK TEST	0.00	00:03



SGT Ralph J Bartlett

08-30-10

WC/0 #162