

SUPERVISOR TEST




-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 09/14/2010

Operator Name: RALPH J BARTLETT  
 Permit #: 26222  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:00
INTERNAL STANDARD	VERIFIED	04:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:01
BLANK TEST	0.00	04:02
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	04:02
BLANK TEST	0.00	04:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:03
BLANK TEST	0.00	04:04



SOT Ralph J Bartlett

WCSD #162

09-19-10