

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 09/20/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10110

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:27
INTERNAL STANDARD	VERIFIED	02:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:27
BLANK TEST	0.00	02:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:28
BLANK TEST	0.00	02:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:30
BLANK TEST	0.00	02:30



015 Ralph Bartlett 09-20-10

WESB #162