

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/29/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:13
INTERNAL STANDARD	VERIFIED	02:13
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:14
BLANK TEST	0.00	02:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:15
BLANK TEST	0.00	02:15
REFERENCE STANDARD TEMP	33.95c	
REFERENCE STANDARD	0.09	02:16
BLANK TEST	0.00	02:16



SGT RALPH BARTLETT WCSO #162 9/29/10