

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/21/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:40
INTERNAL STANDARD	VERIFIED	04:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:41
BLANK TEST	0.00	04:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	04:42
BLANK TEST	0.00	04:43
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	04:43
BLANK TEST	0.00	04:44



SA Ralph Bartlett 10/21/10 WESO #162