

SUPERVISOR TEST




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 10/27/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:38
INTERNAL STANDARD	VERIFIED	01:38
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	01:38
BLANK TEST	0.00	01:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:39
BLANK TEST	0.00	01:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	01:40
BLANK TEST	0.00	01:41



SOS *Ralph Bartlett*

WCSO #112 10/27/10