

SOLUTION CHANGE




WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 11/01/2010

Operator Name: RALPH J BARTLETT  
Permit #: 26222  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:38
INTERNAL STANDARD	VERIFIED	03:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	03:39
BLANK TEST	0.00	03:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:40
BLANK TEST	0.00	03:41



SOS Rupp / Bartlett were SHERID #12 11/01/10