

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/24/2010

Operator Name: CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10140

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:54
INTERNAL STANDARD	VERIFIED	10:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:55
BLANK TEST	0.00	10:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:56
BLANK TEST	0.00	10:57
REFERENCE STANDARD TEMP	33.97c	
REFERENCE STANDARD	0.09	10:57
BLANK TEST	0.00	10:58

Charles B. Geisler #248

