

SOLUTION CHANGE

-----  
 WARREN CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 144506  
 -----

Date: 12/03/2010

Operator Name: CHARLES B GEISLER  
 Permit #: 33757  
 Operator Agency/Dept: WARREN CO. SO  
 Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:49
INTERNAL STANDARD	VERIFIED	13:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:50
BLANK TEST	0.00	13:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:51
BLANK TEST	0.00	13:51
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	13:52
BLANK TEST	0.00	13:52

*Charles B. Geisler #248*

