

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/11/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:06
INTERNAL STANDARD	VERIFIED	02:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:07
BLANK TEST	0.00	02:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:09
BLANK TEST	0.00	02:10



505 *Ralph J Bartlett*

12/11/10

WCSO SHERIFF #162