

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/17/2010

Operator Name: RALPH J BARTLETT
Permit #: 26222
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 10190

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:59
INTERNAL STANDARD	VERIFIED	01:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:59
BLANK TEST	0.00	02:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	02:00
BLANK TEST	0.00	02:01
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	02:01
BLANK TEST	0.00	02:02



Sgt Ralph Bartlett

WCSO #162

12/17/10